

# Certificate of Analysis

RiboGrip® RNase Inhibitor (220 U/μl)

Cat. No.

Lot No.

Pack Lot No.

Expiry date

**Concentration:** 220 U/μl

**Shipping:** at ambient temperature

**Storage and stability\*:**

- Routine storage at –20°C (–28°C to –18°C) until Expiry Date
- Stable at room temperature (25°C) up to 1 month.

**Safety precautions:** Please refer to Safety Data Sheet for more information.

Manufactured by Solis BioDyne in compliance with the ISO 9001 and ISO 13485 certified Quality Management System.

**Quality Control Assays:**

Assay	Result
Protein purity assay (SDS-PAGE)	passed
Inhibition of ribonuclease A	passed
Endodeoxyribonuclease activity	passed
Exodeoxyribonuclease activity	passed
Functionality testing	passed

COA-06-26 v1

Effective from 01.12.2023

*\*Product stability is assessed using routine QC assays and QC criteria set forth in the product specification and are intended to provide guidelines for shipping and storage conditions only. Customer or its designee shall be responsible for conducting all necessary stability testing applicable to their assay and/or QC criteria, and to comply with any applicable regulatory requirements or guidelines. Such stability testing shall include testing to validate the lead times for shipment, the shelf life of, and the product specifications applicable to shipment, storage and handling of the assay assembled and packed by the customer.*

**FOR RESEARCH USE ONLY. NOT FOR USE IN  
DIAGNOSTIC PROCEDURES.**

**APPROVED BY:**

*Head of Quality and Product Management*