

# Certificate of Analysis

## SolisAcura™ Probe Genotyping qPCR Mix (ROX), 5x

**Cat. No.**

**Lot No.**

**Pack Lot No.**

**Expiry date**

**Concentration:** 5x concentrate

**Mix Composition:**

- SolisAcura™ Exo(+) DNA Polymerase
- 5x qPCR buffer
- 15 mM MgCl<sub>2</sub> (1x PCR solution – 3 mM)
- dNTPs
- Passive reference dye based on ROX

**Shipping:** at room temperature

**Storage and stability\*:** Routine storage at -20°C (-28...-18°C) until Expiry Date. Stable at room temperature (15–25°C) for 1 month.

**Freeze-thaw stability:** 10 cycles

**Safety precautions:** Please refer to Safety Data Sheet for more information

Manufactured by Solis BioDyne in compliance with the ISO 9001 and ISO 13485 certified Quality Management System.

**Quality Control Assays:**

SolisAcura™ Probe Genotyping qPCR Mix (ROX):

| Assay                                        | Result |
|----------------------------------------------|--------|
| Amplification efficiency                     | passed |
| qPCR reproducibility                         | passed |
| DNA contamination (human gDNA)               | passed |
| Clustering of samples according to genotypes | passed |

Note - Standard Quality Control Tests are performed for each individual component included in the product and meet the designated specifications.

*\*Product stability is assessed using routine QC assays and QC criteria set forth in the product specification and are intended to provide guidelines for shipping and storage conditions only. Customer or its designee shall be responsible for conducting all necessary stability testing applicable to their assay and/or QC criteria, and to comply with any applicable regulatory requirements or guidelines. Such stability testing shall include testing to validate the lead times for shipment, the shelf life of, and the product specifications applicable to shipment, storage and handling of the assay assembled and packed by the customer.*

COA-38-02 v1

Effective from 06.10.2025

**FOR RESEARCH USE ONLY. NOT FOR USE IN  
DIAGNOSTIC PROCEDURES.**

**APPROVED BY:**

*Head of Quality and Product Management*