

SOLIScript® 1-step Multiplex Probe Kit (ROX)

| Catalogue Number | Pack Size |
|------------------|-----------------------|
| 08-59-0000S | 50 x 20 µl reactions |
| 08-59-00250 | 250 x 20 µl reactions |



Shipping:

At room temperature

Batch Number and Expiry Date:

See vial

Storage and Stability:

- Routine storage at -20°C until expiry date
- Can be stored at +4°C for up to 6 months
- Stability at room temperature (15–25°C) for 1 month
- Freeze-thaw stability: 30 cycles
- Store 5x One-step Multiplex Probe Mix (ROX) protected from light

Stability at room
temperature
1 MONTH

Store at -20°C

Reaction setup:

At room temperature

Manufactured by Solis BioDyne, in compliance with the ISO 9001 and ISO 13485 certified Quality Management System.

DS-08-59 v2

1/8

Content:

| Component | Catalogue Number | |
|--|-------------------------------|--------------------------------|
| | 08-55-0000S 50 x 20 µl rxn | 08-55-00250 250 x 20 µl rxn |
| 40x One-step SOLIScript® Mix | 25 µl | 125 µl |
| 5x One-step Multiplex Probe Mix (ROX) | 200 µl | 1000 µl |
| Water, nuclease-free | 1.25 ml | 5.0 ml |

Note: To avoid repeated freezing and thawing as well as to minimize the contamination risk of stock solutions of reagents, it is highly recommended to divide large-volume stocks into several smaller aliquots and store them at -20°C.

Mix compositions:

| Kit component | Description |
|--|---|
| 40x One-step SOLIScript® Mix | SOLIScript® Reverse Transcriptase, RiboGrip™ RNase Inhibitor |
| 5x One-step Multiplex Probe Mix (ROX) | HOT FIREPol® DNA polymerase, dNTPs (including dUTPs), passive reference dye based on ROX dye, 15 mM MgCl ₂ (1x RT-qPCR solution – 3 mM MgCl ₂) |

Note: Heat-inactivated UNG is not included in the Kit and should be purchased separately from other supplier.

Compatible real-time instruments:

qPCR cyclers where ROX is used as a passive reference dye for normalization of fluorescent signal.

3/8

Product description:

- SOLIScript® 1-step Multiplex Probe Kit (ROX) is optimized for sensitive and accurate quantification of RNA targets by real-time one-step RT-PCR (RT-qPCR) using dual-labeled hydrolysis probes (e.g. TaqMan® probes), and is suitable for detection and quantification of up to four targets simultaneously.
- The Kit comes in 3 tubes (including water) and contains all the components necessary (except RNA template, probe(s), and primers) to perform RT-qPCR in a single tube on qPCR cyclers that require an internal reference dye ROX for normalization of fluorescent signal.
- 40x One-step SOLIScript® Mix comprises *in silico*-engineered SOLIScript® Reverse Transcriptase and RiboGrip™ RNase Inhibitor. SOLIScript® is a thermostable reverse transcriptase active at temperatures up to 60°C, beneficial when using templates with high levels of secondary structure. RiboGrip™ inactivates RNase A to protect RNA sample from degradation and increase cDNA yield in reactions with low RNA amounts.
- 5x One-step Multiplex Probe Mix (ROX) contains HOT FIREPol® DNA Polymerase in a unique reaction buffer that maximizes the performance of the Kit with high GC-content regions. The buffer includes dUTPs to prevent carry-over contamination from previous amplifications if used in combination with heat-inactivated uracil-N-glycosylase (UNG).
- The Kit can be used for gene expression analysis and low-copy gene detection, gene knockdown validation, miRNA profiling and quantification, characterization of GMOs, RNA viral pathogen detection and quantification.

2/8

Step-by-step guidelines:

1. Thaw the reagents at room temperature. Mix each reagent by gentle vortexing or pipetting up and down, then centrifuge briefly. **NOTE:** Minimize the exposure of the 5x One-step Multiplex Probe Mix (ROX) and fluorescent probe(s) to light.
2. Prepare a reaction mix. Add all components except the template RNA. **NOTE:** Include no-template (NTC) and no-RT (NRT) controls. In NTC replace RNA, in NRT replace 40x One-step SOLIScript® Mix with corresponding amounts of nuclease-free water. If combining different assays on a single plate, include at least one NRT for each assay for every RNA.

| Component | Volume ¹ | Final conc. |
|---------------------------------------|---------------------|-----------------------|
| 40x One-step SOLIScript® Mix | 0.5 µl | 1x |
| 5x One-step Multiplex Probe Mix (ROX) | 4 µl | 1x |
| Forward Primer (10 µM) ² | 0.8 µl | 400 nM |
| Reverse Primer (10 µM) ² | 0.8 µl | 400 nM |
| Probe ² | Variable | 250 nM |
| Template RNA ² | Variable | 1 pg–1 µg (total RNA) |
| Water, nuclease-free | up to 20 µl | |
| Total reaction volume | 20 µl | |

¹ Reaction volume can be scaled from 10 to 50 µl. Scale components proportionally according to sample number and reaction volumes. Use enough of each reagent for your reactions, plus 5–10% extra volume to accommodate pipetting errors.

² See page 6–7 for Recommendations on primers/probe design, template preparation and additional guidelines.

4/8

- Mix the reaction mix thoroughly, then centrifuge briefly. Dispense appropriate volumes of mix into PCR wells.
- Add template RNA to the PCR wells. Seal the wells using the procedure recommended for the cycling instrument being used, and centrifuge the reactions briefly.
- Run your RT-qPCR assay as shown in table below. **NOTE:** When programming your RT-qPCR run, choose the detection channel(s) of the qPCR instrument that correspond to the fluorophore label(s) of the target-specific probe(s) present in the assay. Acquisition of real-time data generated by fluorogenic probe(s) should be performed as recommended by the instrument's manufacturer.

| | | Temperature | Time |
|------------------------------------|---------------------|-------------|--------|
| Reverse transcription ¹ | | 50°C | 15 min |
| Enzyme activation ² | | 95°C | 10 min |
| 40 cycles | Denaturation | 95°C | 15 sec |
| | Annealing/Extension | 60°C | 60 sec |

¹ A 50°C RT step temperature is optimal for SOLIScript® Reverse Transcriptase. SOLIScript® Reverse Transcriptase is active at 45–60°C. For difficult templates with high secondary structure, the temperature may be increased to 60°C.

² IMPORTANT: Enzyme activation step at 95°C for 10 min is crucial for full activation of DNA polymerase and inactivation of reverse transcriptase.

- Collect and analyze the data according to the instrument-specific instructions. Verify the amplification curve, set up a standard curve if absolute quantification is desired.

5/8

- Suitable primers are typically 18–24 bp long, with GC-content 35–65 %, and target a 50–150 bp amplicon. The optimal melting temperature (T_m) of the primers is 58–60°C. Primers with T_m outside of this range may and/or amplicons longer than 150 bp require optimization of cycling conditions. For multiplex reactions, design primer pairs with similar annealing temperatures for all targets to be amplified. Self-complementarity of primers should be avoided.
- Optimal results may require titration of primer concentration between 200 and 600 nM. A final concentration of 400 nM each primer is suitable for most applications.

Probe:

- Fluorescent reporter dye(s) and quencher(s) for your RT-qPCR analysis should be chosen according to the detection capabilities of your qPCR platform. Refer to your instrument manufacturer's guidelines for information specific to your particular instrument.
- The length of the probe is 9–40 bases (hydrolysis probes: up to 30 bases). The probe binding site should be in close proximity to the forward or reverse primer; however, binding sites should not overlap. Probe can bind to either strand of the target.
- T_m of the probe should be 6–8°C higher than the T_m of the primers. In case the T_m of the probe is not optimal, less of the probe is bound to the target sequence. This may result in compromised sensitivity, reduced fluorescence signal and not a true representation of the target amount in the sample.
- Optimal results may require titration of probe concentration between 100 and 250 nM. A final concentration of each 250 nM probe is suitable for most applications.

7/8

Recommendations for a successful RT-qPCR experiment

Template:

- Ratios of the absorbance at 230, 260, and 280 nm are used to assess the purity of nucleic acid samples. "Good RNA" has 260/280 and 260/230 ratios equal or higher than 1.8 and 2.0, respectively. The presence of contaminants such as phenol, carbohydrates or EDTA may compromise the results.
- Treat your purified RNA sample with RNase-free DNase I to remove contaminating genomic DNA which can act as template during reaction.
- Starting material can range from 1 pg to 1 µg of purified total RNA per 20 µl reaction. To determine the optimal template quantity (especially when amplifying more than one target and/or using other types of input RNA, e.g. mRNA or viral RNA), perform your reactions on a serially diluted template (e.g. 10-fold dilution series). Select the quantity that produces the earliest Ct without inhibiting your amplification.

Primers:

- Use dedicated software, such as open source Primer3 and NCBI Primer-BLAST to design target-specific primers. To prevent amplification from genomic DNA, design primers spanning exon-exon junctions of the target mRNA, or primers that hybridize with sequences in consecutive exons flanking a long (e.g. 1 kb) intron. Amplification of more than one target in a single reaction may require considerable optimization, individual assays must be validated prior to multiplexing.

6/8

Safety precautions:

Please refer to Safety Data Sheet for more information.

Technical support:

Contact your sales representative for any questions or send an email to support@solisbioDYNE.com

DS-08-59 v2

Revised 08.06.2020

This product is supplied for research use only. It is suitable for use as a component of molecular diagnostic assays, where applicable country laws allow. This product alone does not provide any diagnostic result. Covered by patent EP2501716, made following the methods of US Patent No 9,321,999.

RiboGrip™ is a trademark of Solis BioDYNE OÜ. FIREPoI® and SOLIScript® are registered trademarks of Solis BioDYNE OÜ. TaqMan® is a registered trademark of Roche Molecular Systems, Inc.



8/8